



At QCS we know how important it is to keep everyone safe, so we want to ensure you are aware of a product recall.

Type IIR masks Cardinal Health Product Notice – all lots

Although these masks meet the breathability, filtration and splash resistance requirements of BS EN 14683, in light of ongoing monitoring, further complaints reported and testing from the manufacturer on the masks, the Medicines and Healthcare Regulatory Agency (MHRA) recommends that all lots of this product are disposed of locally. There is a risk to staff wearing the mask if the foam strip on the mask flakes and enters their airway or mouth.

There have also been complaints of the ties and/or stitching coming away from the mask.

Product Affected:

There is one product impacted

- ✓ Mask, Type IIR
- ✓ Supplier: Cardinal Health
- ✓ NPC: BWM028
- ✓ MPC: AT74535UK

Applies to all lots of AT74535UK



Affected boxes

- ✿ Boxes have been labelled with a new expiry date because they were subject to shelf life extension testing by the manufacturer in 2013/2014 and passed several relevant tests to support a new expiry date
- ✿ When potential issues were identified in May, a sample size of 7 lots was put through
- ✿ Further testing by the manufacturer in June 2020. 6 of the 7 lots did not pass a material inspection of the foam strip
- ✿ However, these masks meet the breathability, filtration and splash resistance requirements of BS EN 14683
- ✿ MHRA has recommended that all lots of this product are disposed of locally

What you need to do now:

- ✓ Identify, stop using and immediately quarantine all lots relating to BWM028 of the Cardinal Type IIR masks
- ✓ All stock should be disposed of locally
- ✓ If you have a shortage of Type IIR masks, please notify NSDR on 0800 915 9964 who will arrange an emergency delivery if required
- ✓ For all general enquiries for the PPE Dedicated Channel, call Customer Services on 0800 876 6802